

1

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 1

1

Lot Number:



Exp. Date: 2013-04

Parameter	Units	Expected Range		+/- 3 SD Expected Range	
		Lower Limit	Upper Limit	Lower Limit	Upper Limit
pH	-	 6.632	 6.732	 6.607	 6.757
PO ₂	mm Hg	 60.9	 73.3	 57.8	 76.4
PCO ₂	mm Hg	 58.4	 72.0	 55.0	 75.4
Gluc	g/L	 0.75	 1.25	 0.62	 1.38
Lac	g/L	 0.78	 1.28	 0.65	 1.41
NH ₄ ⁺	mmol/L	 0.61	 1.11	 0.48	 1.24
Na ⁺	mmol/L	 69.7	 85.3		
K ⁺	mmol/L	 1.71	 2.29		
Ca ⁺⁺	mmol/L	 0.42	 0.66		
PO ₄	mmol/L	 3.56	 4.76		
p	mg/L	 110.4	 147.6		
Osmolality	mOsm/kg H ₂ O	 214	 224		