










4

Assay Data Sheet - Expected Ranges BioProfile® FLEX Controls - LEVEL 4

4

Lot Number:  209332		Exp. Date: 2013-12	
Expected Range			
Parameter	Units	Lower Limit	Upper Limit
Gln	mmol/L	 0.71	 1.21
Glu	mmol/L	 0.26	 0.76
Gluc	g/L	 0.69	 1.19
Lac	g/L	 0.77	 1.27

nova[®]
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08242B 2008-01