

5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  208335		Exp. Date: 2014-02	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 4.10	 5.60
Glu	mmol/L	 1.70	 2.30
Gluc	g/L	 8.08	 12.08
Lac	g/L	 4.30	 5.80



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800