


































3

Assay Data Sheet - Expected Ranges BioProfile® FLEX Controls - LEVEL 3

3

Lot Number:  900708				Exp. Date: 2020-07	
Expected Range				+/- 3 SD Expected Range	
Parameter	Units	Lower Limit	Upper Limit	Lower Limit	Upper Limit
pH	-	 7.345	 7.445	 7.320	 7.470
PO ₂	mm Hg	 133.1	 161.7	 126.0	 168.8
PCO ₂	mm Hg	 13.0	 19.4	 11.4	 21.0
Gluc	g/L	 12.75	 17.25	 11.62	 18.38
Lac	g/L	 4.25	 5.75	 3.87	 6.13
NH ₄ ⁺	mmol/L	 8.24	 11.24	 7.49	 11.99
Na ⁺	mmol/L	 166.7	 203.3		
K ⁺	mmol/L	 5.55	 6.85		
Ca ⁺⁺	mmol/L	 2.25	 2.85		
Osmolality	mOsm/kg H ₂ O	 605	 621		

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08241A 2007-08