

5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  905150		Exp. Date: 2020-08	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 4.46	 5.96
Glu	mmol/L	 1.81	 2.41
Gluc	g/L	 8.19	 12.19
Lac	g/L	 4.46	 5.96



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800