

**1**

**Assay Data Sheet - Expected Ranges  
BioProfile® FLEX Controls - LEVEL 1**

**1**

608870

Lot Number:

Exp. Date: 2017-09

Parameter	Units	Expected Range		+/- 3 SD Expected Range	
		Lower Limit	Upper Limit	Lower Limit	Upper Limit
pH	-	6.620	6.720	6.595	6.745
PO <sub>2</sub>	mm Hg	58.5	70.9	55.4	74.0
PCO <sub>2</sub>	mm Hg	60.9	74.5	57.5	77.9
Gluc	g/L	0.75	1.25	0.62	1.38
Lac	g/L	0.78	1.28	0.65	1.41
NH <sub>4</sub> <sup>+</sup>	mmol/L	0.61	1.11	0.48	1.24
Na <sup>+</sup>	mmol/L	69.7	85.3		
K <sup>+</sup>	mmol/L	1.71	2.29		
Ca <sup>++</sup>	mmol/L	0.50	0.74		
PO <sub>4</sub>	mmol/L	3.45	4.65		
p	mg/L	107.0	144.2		
Osmolality	mOsm/kg H <sub>2</sub> O	217	227		

*NOVA*  
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08239B 2011-03