






































**2**

Assay Data Sheet - Expected Ranges  
 BioProfile® FLEX Controls - LEVEL 2

**2**

Lot Number:  601145		Exp. Date: 2017-07	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
pH	-	 6.891	 6.991
PO <sub>2</sub>	mm Hg	 101.0	 121.0
PCO <sub>2</sub>	mm Hg	 41.3	 50.1
Gluc	g/L	 4.25	 5.75
Lac	g/L	 2.55	 3.45
NH <sub>4</sub> <sup>+</sup>	mmol/L	 3.85	 5.65
Na <sup>+</sup>	mmol/L	 122.5	 140.9
K <sup>+</sup>	mmol/L	 3.58	 4.42
Ca <sup>++</sup>	mmol/L	 0.67	 0.93
PO <sub>4</sub>	mmol/L	 0.72	 1.12
P	mg/L	 22.3	 34.7
Osmolality	mOsm/kg H <sub>2</sub> O	 396	 408
		+/- 3 SD Expected Range	
		Lower Limit	Upper Limit
		 6.866	 7.016
		 96.0	 126.0
		 39.2	 52.2
		 3.87	 6.13
		 2.32	 3.68
		 3.35	 6.15



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08240B 2011-03