






































2

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 2

2

Lot Number:  616234		Exp. Date: 2017-12	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
pH	-	 6.900	 7.000
PO ₂	mm Hg	 97.3	 117.3
PCO ₂	mm Hg	 38.4	 47.2
Gluc	g/L	 4.25	 5.75
Lac	g/L	 2.55	 3.45
NH ₄ ⁺	mmol/L	 3.85	 5.65
Na ⁺	mmol/L	 122.5	 140.9
K ⁺	mmol/L	 3.58	 4.42
Ca ⁺⁺	mmol/L	 0.67	 0.93
PO ₄	mmol/L	 0.72	 1.12
P	mg/L	 22.3	 34.7
Osmolality	mOsm/kg H ₂ O	 397	 409

Exp. Date: 2017-12	
+/- 3 SD Expected Range	
Lower Limit	Upper Limit
 6.875	 7.025
 92.3	 122.3
 36.3	 49.3
 3.87	 6.13
 2.32	 3.68
 3.35	 6.15

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08240B 2011-03