


































3

Assay Data Sheet - Expected Ranges BioProfile® FLEX Controls - LEVEL 3

3

Lot Number:  135723				Exp. Date: 2022-12	
Expected Range				+/- 3 SD Expected Range	
Parameter	Units	Lower Limit	Upper Limit	Lower Limit	Upper Limit
pH	-	 7.307	 7.407	 7.282	 7.432
PO ₂	mm Hg	 131.5	 160.1	 124.4	 167.2
PCO ₂	mm Hg	 16.1	 22.5	 14.5	 24.1
Gluc	g/L	 12.75	 17.25	 11.62	 18.38
Lac	g/L	 4.25	 5.75	 3.87	 6.13
NH ₄ ⁺	mmol/L	 8.24	 11.24	 7.49	 11.99
Na ⁺	mmol/L	 166.7	 203.3		
K ⁺	mmol/L	 5.55	 6.85		
Ca ⁺⁺	mmol/L	 2.25	 2.85		
Osmolality	mOsm/kg H ₂ O	 602	 618		

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08241A 2007-08