


































3

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 3

3

Lot Number:  632745		Exp. Date: 2018-05	
Expected Range			
Parameter	Units	Lower Limit	Upper Limit
		+/- 3 SD Lower Limit	Expected Range Upper Limit
pH	-	 7.333	 7.433
PO ₂	mm Hg	 137.2	 165.8
PCO ₂	mm Hg	 14.9	 21.3
Gluc	g/L	 12.75	 17.25
Lac	g/L	 4.25	 5.75
NH ₄ ⁺	mmol/L	 8.24	 11.24
Na ⁺	mmol/L	 166.7	 203.3
K ⁺	mmol/L	 5.55	 6.85
Ca ⁺⁺	mmol/L	 2.25	 2.85
Osmolality	mOsm/kg H ₂ O	 605	 621
		 7.308	 7.458
		 130.1	 172.9
		 13.3	 22.9
		 11.62	 18.38
		 3.87	 6.13
		 7.49	 11.99



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08241A 2007-08