


































**3**

**Assay Data Sheet - Expected Ranges  
BioProfile® FLEX Controls - LEVEL 3**

**3**

<b>Lot Number:</b>  706945		<b>Exp. Date:</b> 2018-09	
<b>Parameter</b>	<b>Units</b>	<b>Expected Range</b>	
		<b>Lower Limit</b>	<b>Upper Limit</b>
<b>pH</b>	-	 7.343	 7.443
		 7.318	 7.468
<b>PO<sub>2</sub></b>	mm Hg	 136.2	 164.8
		 129.1	 171.9
<b>PCO<sub>2</sub></b>	mm Hg	 13.2	 19.6
		 11.6	 21.2
<b>Gluc</b>	g/L	 12.75	 17.25
		 11.62	 18.38
<b>Lac</b>	g/L	 4.25	 5.75
		 3.87	 6.13
<b>NH<sub>4</sub><sup>+</sup></b>	mmol/L	 8.24	 11.24
		 7.49	 11.99
<b>Na<sup>+</sup></b>	mmol/L	 166.7	 203.3
<b>K<sup>+</sup></b>	mmol/L	 5.55	 6.85
<b>Ca<sup>++</sup></b>	mmol/L	 2.25	 2.85
<b>Osmolality</b>	mOsm/kg H <sub>2</sub> O	 604	 620

***NOVA***  
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800