










4

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 4

4

Lot Number:  705206		Exp. Date: 2018-08	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 0.74	 1.24
Glu	mmol/L	 0.26	 0.76
Gluc	g/L	 0.74	 1.24
Lac	g/L	 0.74	 1.24



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800