









5

Assay Data Sheet - Expected Ranges  
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:   
005123

Exp. Date: 2021-08

Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 4.46	 5.96
Glu	mmol/L	 1.79	 2.39
Gluc	g/L	 8.20	 12.20
Lac	g/L	 4.51	 6.01

*NOVA*  
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800