


5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  134713		Exp. Date: 2023-06	
Expected Range			
Parameter	Units	Lower Limit	Upper Limit
Gln	mmol/L	 4.43	 5.93
Glu	mmol/L	 1.77	 2.37
Gluc	g/L	 8.22	 12.22
Lac	g/L	 4.42	 5.92

nova
biomedical
200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800