


























2

Assay Data Sheet - Expected Ranges
 BioProfile® FLEX Controls - LEVEL 2

2

Lot Number:  623035		Exp. Date: 2018-02	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
pH	-	 6.905	 7.005
PO ₂	mm Hg	 99.0	 119.0
PCO ₂	mm Hg	 35.9	 44.7
Gluc	g/L	 4.25	 5.75
Lac	g/L	 2.55	 3.45
NH ₄ ⁺	mmol/L	 3.85	 5.65
Na ⁺	mmol/L	 122.5	 140.9
K ⁺	mmol/L	 3.58	 4.42
Ca ⁺⁺	mmol/L	 0.67	 0.93
PO ₄	mmol/L	 0.72	 1.12
P	mg/L	 22.3	 34.7
Osmolality	mOsm/kg H ₂ O	 396	 408

Exp. Date: 2018-02

+/- 3 SD Expected Range

Lower Limit

Upper Limit

7.030

124.0

46.8

6.13

3.68

6.15

140.9


4.42

0.93

1.12

34.7

408



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08240B 2011-03