


































3

**Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 3**

3

Lot Number:  535507		Exp. Date: 2017-06	
		+/- 3 SD Expected Range	
Parameter	Units	Lower Limit	Upper Limit
pH	-	 7.321	 7.421
PO ₂	mm Hg	 136.5	 165.1
PCO ₂	mm Hg	 14.2	 20.6
Gluc	g/L	 12.75	 17.25
Lac	g/L	 4.25	 5.75
NH ₄ ⁺	mmol/L	 8.24	 11.24
Na ⁺	mmol/L	 166.7	 203.3
K ⁺	mmol/L	 5.55	 6.85
Ca ⁺⁺	mmol/L	 2.25	 2.85
Osmolality	mOsm/kg H ₂ O	 605	 621
		Lower Limit	Upper Limit
		 7.296	 7.446
		 129.4	 172.2
		 12.6	 22.2
		 11.62	 18.38
		 3.87	 6.13
		 7.49	 11.99

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800

LFN 08241A 2007-08